

CERTIFICATE

Biocompatibility Test

Material tested

ZENO Al_{Crown}

Dental Ceramic, ISO 6872

Composition

Al₂O₃
> 99,5%

w/w %

Manufacturer

WIELAND Dental + Technik GmbH & Co. KG
Schwenninger Straße 13 · 75179 Pforzheim, Germany

Tests

We confirm that the following tests for determining the biocompatibility were carried out in accordance with the international standards ISO 10993, "Biological evaluation of medical devices" (EN ISO 10993-1, EN ISO 10993-2, EN ISO 10993-5, EN ISO 10993-12) and EN ISO 7405, "Dentistry – Preclinical evaluation of biocompatibility of medical devices used in dentistry – Test methods for dental materials".

The tests were performed according to the OECD directives "Good Laboratory Practice" (GLP) by the Institute BSL Bioservice Scientific Laboratories.

The test items were produced with the ZENO CAD/CAM System by the dental laboratory of WIELAND Dental + Technik.

Cytotoxicity

The cytotoxic potential of the dental ceramic was tested in vitro with L-929 fibroblasts
Method: "Test on extracts", XTT staining, EN ISO 10993-5: 1999, EN ISO 10993-12: 2007 and EN ISO 7405: 1997 (5.4.a)3).

Test result:

ZENO Al_{Crown} had no cytotoxic potential

Dr. G. Henning, Dental Engineering
Unterer Lichsenweg 13 · D-79541 Lörrach



Lörrach, 2008-04-02